



## Certificate of Analysis

CytoFix® Low            CF04L  
Batch Number            28088  
Expiry Date              22<sup>nd</sup> December 2010

Expected Values						
	% Positive	Range		Absolute (/ul)	Range (/ul)	
		min	max		min	max
CD3+	71	69	73	1150	999	1301
CD3+CD4+	15	13	18	246	188	303
CD3+ CD8+	49	45	52	782	650	914
CD16+56+	13	11	15	214	160	269
CD19+	16	13	18	255	203	307

Lymphocytes	16	11	21	N/A	N/A	N/A
Monocytes	6	1	10	N/A	N/A	N/A
Granulocytes	54	44	64	N/A	N/A	N/A


These values have been determined using Becton Dickenson MultiTest reagents with TruCount tubes using the no wash method

CytoFix® Normal        CF04N  
Batch Number            28089  
Expiry Date              22<sup>nd</sup> December 2010

Expected Values						
	% Positive	Range		Absolute (/ul)	Range (/ul)	
		min	max		min	max
CD3+	82	80	84	1201	1043	1359
CD3+CD4+	44	41	47	642	534	750
CD3+ CD8+	35	32	38	505	403	607
CD16+56+	9	7	11	128	93	164
CD19+	8	7	10	125	93	157

Lymphocytes	20	16	24	N/A	N/A	N/A
Monocytes	7	5	9	N/A	N/A	N/A
Granulocytes	48	42	54	N/A	N/A	N/A

These values have been determined using Becton Dickenson MultiTest reagents with TruCount tubes using the no wash method

  
-----  
Tim Almond,  
CEO, Cytomark Ltd