

CD4 Blood Controls Certificate of Analysis

CD4 Blood Control

Low CF04L
Batch Number 29208
Expiry Date 07 August 2014

Expected Values						
	% Positive	Range (%)		Absolute (/μl)	Range (/μl)	
		min	max		min	max
CD3+	58.4	56.2	60.6	1004	899	1109
CD3+ CD4+	13.8	12.3	15.4	238	214	261
CD3+ CD8+	31.0	29.5	32.4	533	467	599
NK Cells	24.6	23.1	26.0	423	368	477
B Cells	14.6	12.9	16.3	251	202	300

Lymphocytes	19.0	17.6	20.4	N/A	N/A	N/A
Monocytes	0.9	0.7	1.1	N/A	N/A	N/A
Granulocytes	35.6	32.5	38.6	N/A	N/A	N/A

These values have been determined using Cytognos reagents using the lyse, no wash method.

CD4 Blood Control

Normal CF04N
Batch Number 29209
Expiry Date 07 August 2014

Expected Values						
	% Positive	Range (%)		Absolute (/μl)	Range (/μl)	
		min	max		min	max
CD3+	75.6	73.1	78.0	995	929	1060
CD3+ CD4+	37.2	34.1	40.2	489	437	542
CD3+ CD8+	23.2	22.0	24.5	306	282	330
NK Cells	12.9	11.9	13.8	169	148	191
B Cells	10.2	8.2	12.2	134	106	162

Lymphocytes	18.2	16.6	19.8	N/A	N/A	N/A
Monocytes	1.6	1.3	1.9	N/A	N/A	N/A
Granulocytes	29.3	27.0	31.5	N/A	N/A	N/A

These values have been determined using Cytognos reagents using the lyse, no wash method.



Quality Manager

DANIEL HARRISON

28 MAY 2014