

# TransFix<sup>®</sup>

## Certificate of Analysis TransFix<sup>®</sup>/EDTA Vacuum Blood Collection Tubes (TVT)



**LOT** : 29392

 : 2017-05

 Start Date of Manufacture: **2015-10-29**

 TransFix Batch Number used in TVT manufacture: **29342**

 Start Date of Quality Control Testing: **2015-11-20**

	A	B		
	TVT Day 0	TVT Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	74.7	74.8	0.1	PASS
% CD4+	29.5	28.4	3.6	PASS
CD3+ Abs Cnt	1538	1366	11.2	PASS
CD4+ Abs Cnt	607	519	14.4	PASS
CD8+ Abs Cnt	526	451	14.1	PASS
CD45+ Abs Cnt	2058	1827	11.2	PASS

**Pass Criterion:**

The pass criterion is a less than 15% difference between the lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

**Comments:**

5 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager



Date: 2015-12-02