

# TransFix®

## Certificate of Analysis TransFix® Cellular Antigen Stabilisation Reagent



**LOT** : 29545



: 2019-02-16

Start Date of Manufacture: **2017-02-16**

Start Date of Quality Control Testing: **2017-02-24**

	A	B		
	TransFix Day 0	TransFix Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	76.3	77.0	0.9	PASS
% CD4+	63.5	67.1	5.8	PASS
CD3+ Abs Cnt	1276	1237	3.0	PASS
CD4+ Abs Cnt	809	830	2.6	PASS
CD8+ Abs Cnt	303	314	3.9	PASS
CD45+ Abs Cnt	1673	1606	4.0	PASS

### Pass Criterion:

The pass criterion is a less than 15% difference between the lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

### Comments:

3 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager



Date: 2017-03-07

