

TransFix[®]

Certificate of Analysis TransFix[®] Cellular Antigen Stabilisation Reagent



LOT : 29571

 : 2019-05-17

 Start Date of Manufacture: **2017-05-17**

 Start Date of Quality Control Testing: **2017-06-30**

	A	B		
	Control Day 0	TransFix Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	73.3	74.8	2.1	PASS
% CD4+	63.6	60.9	4.2	PASS
CD3+ Abs Cnt	1851	1714	7.4	PASS
CD4+ Abs Cnt	1176	1044	11.2	PASS
CD8+ Abs Cnt	504	458	9.1	PASS
CD45+ Abs Cnt	2525	2291	9.3	PASS

Pass Criterion:

The pass criterion is a less than 15% difference between the lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

Comments:

3 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager



Date: 2017-07-12