

TransFix®

Certificate of Analysis TransFix®/EDTA CSF Sample Storage Tubes

RUO

LOT : 29536

 : 2018-01-19

Start Date of Manufacture: **2017-01-19**

Start Date of Quality Control Testing: **2017-01-27**

	A	B		
	Control Day 0	TransFix/EDTA Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	81.8	80.3	1.8	PASS
% CD4+	34.3	33.9	1.1	PASS
CD3+ Abs Cnt	1533	1366	9.2	PASS
CD4+ Abs Cnt	526	463	10.9	PASS
CD8+ Abs Cnt	824	730	11.9	PASS
CD45+ Abs Cnt	1874	1701	11.4	PASS

Pass Criterion:

The pass criterion is a less than 15% difference between the lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

Comments:

3 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager



Date: 2017-02-08

