

TransFix®

Certificate of Analysis TransFix®/EDTA CSF Sample Storage Tubes

RUO

LOT : 29555

 : 2018-03-31

Start Date of Manufacture: **2017-03-16**

Start Date of Quality Control Testing: **2017-04-07**

	A	B		
	Control Day 0	TransFix/EDTA Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	78.9	74.3	5.8	PASS
% CD4+	43.3	40.1	7.5	PASS
CD3+ Abs Cnt	1190	1184	0.5	PASS
CD4+ Abs Cnt	653	638	2.3	PASS
CD8+ Abs Cnt	388	373	3.9	PASS
CD45+ Abs Cnt	1508	1592	5.6	PASS

Pass Criterion:

The pass criterion is a less than 15% difference between the lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

Comments:

3 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager



Date: 2017-04-18

