

CD4 Blood Controls Certificate of Analysis

CD4 Blood Control

Low CF04L
Batch Number 29327
Expiry Date 27 July 2015

Expected Values						
	% Positive	Range (%)		Absolute (/μl)	Range (/μl)	
		min	max		min	max
CD3+	60.6	57.7	63.5	780	608	952
CD3+ CD4+	21.1	19.2	23.0	272	201	342
CD3+ CD8+	33.7	32.1	35.3	434	341	526
NK Cells	17.3	15.5	19.2	223	179	266
B Cells	21.5	19.1	23.8	275	237	314
Lymphocytes	16.8	15.3	18.2	N/A	N/A	N/A
Monocytes	1.4	0.6	2.2	N/A	N/A	N/A
Granulocytes	30.8	24.1	37.5	N/A	N/A	N/A

These values have been determined using Cytognos reagents using the lyse, no wash method.

CD4 Blood Control

Normal CF04N
Batch Number 29326
Expiry Date 27 July 2015

Expected Values						
	% Positive	Range (%)		Absolute (/μl)	Range (/μl)	
		min	max		min	max
CD3+	79.3	77.2	81.5	1284	1121	1446
CD3+ CD4+	45.2	43.3	47.1	731	628	834
CD3+ CD8+	28.9	27.7	30.0	467	403	531
NK Cells	8.3	6.6	9.9	134	99	169
B Cells	12.0	11.0	13.0	194	161	227
Lymphocytes	21.5	20.1	22.9	N/A	N/A	N/A
Monocytes	2.9	2.2	3.7	N/A	N/A	N/A
Granulocytes	25.3	21.7	28.9	N/A	N/A	N/A

These values have been determined using Cytognos reagents using the lyse, no wash method.



QUALITY MANAGER
07 MAY 2015