

CD4 Blood Controls Certificate of Analysis

CD4 Blood Control

Low CF04L
 Batch Number 29261
 Expiry Date 06 Jan 2015

Expected Values						
	% Positive	Range (%)		Absolute (/μl)	Range (/μl)	
		min	max		min	max
CD3+	41.8	39.4	44.2	805	675	934
CD3+ CD4+	12.8	11.5	14.1	246	210	281
CD3+ CD8+	21.0	19.3	22.6	404	327	482
NK Cells	31.9	29.0	34.8	614	518	711
B Cells	24.7	21.8	27.6	476	355	596

Lymphocytes	25.4	19.5	31.2	N/A	N/A	N/A
Monocytes	1.4	0.9	1.9	N/A	N/A	N/A
Granulocytes	20.5	17.0	24.0	N/A	N/A	N/A

These values have been determined using Cytognos reagents using the lyse, no wash method.

CD4 Blood Control

Normal CF04N
 Batch Number 29262
 Expiry Date 06 Jan 2015

Expected Values						
	% Positive	Range (%)		Absolute (/μl)	Range (/μl)	
		min	max		min	max
CD3+	78.7	75.6	81.7	1909	1445	2372
CD3+ CD4+	49.7	46.8	52.6	1205	921	1490
CD3+ CD8+	22.6	20.1	25.1	549	391	707
NK Cells	9.9	7.7	12.1	240	182	299
B Cells	10.6	9.2	12.0	258	197	318

Lymphocytes	20.4	17.5	23.3	N/A	N/A	N/A
Monocytes	2.1	0.8	3.4	N/A	N/A	N/A
Granulocytes	37.1	32.3	41.8	N/A	N/A	N/A

These values have been determined using Cytognos reagents using the lyse, no wash method.

QUALITY MANAGER



16 OCT 2014