

# TransFix<sup>®</sup>

## Certificate of Analysis TransFix<sup>®</sup> Cellular Antigen Stabilisation Reagent



**LOT** : 29755

 : 2020-10-15

Start Date of Manufacture: **2018-10-15**

Start Date of Quality Control Testing: **2018-11-08**

	A	B		
	Control Day 0	TransFix Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	79.8	80.9	1.5	Pass
% CD4+	36.9	37.7	2.1	Pass
CD3+ Abs Cnt	1040	1085	4.3	Pass
CD4+ Abs Cnt	384	409	6.5	Pass
CD8+ Abs Cnt	516	541	4.8	Pass
CD45+ Abs Cnt	1304	1340	2.8	Pass

### Pass Criterion:

The pass criterion is a less than 15% difference between the lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

### Comments:

3 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager



Date: 27 Nov 2018

