

TransFix®

Certificate of Analysis TransFix® Cellular Antigen Stabilization Reagent

REF : TF-5-2, TF-5-10, TF-5-25, and TF-5-50

LOT : 29935.60

 : 2023-04-30

Start Date of Manufacture: **2021-06-21**

Start Date of Quality Control Testing: **2021-07-08**

	A	B		
	Control Day 0	TransFix Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	76.7	77.6	1.3	PASS
% CD4+	67.3	67.9	0.8	PASS
CD3+ Abs Cnt	1212	1188	2.0	PASS
CD4+ Abs Cnt	816	807	1.2	PASS
CD8+ Abs Cnt	356	357	0.1	PASS
CD45+ Abs Cnt	1618	1552	4.1	PASS

Pass Criterion:

The pass criterion is a less than 15% difference between the blood lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

Comments:

3 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager



Date: 21 Apr 2022