

# TransFix<sup>®</sup>

## Certificate of Analysis

### TransFix<sup>®</sup>/EDTA Vacuum Blood Collection Tubes (TVT)



 : TVT-03-1, TVT-03-2, TVT-03-50
  : 29957

 : 2023-07-31

Start Date of Manufacture: **2022-01-20**

TransFix Batch Number used in TVT manufacture: **29946**

Start Date of Quality Control Testing: **2020-03-03**

	A	B		
	Control Day 0	TVT Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	73.7	73.3	0.5	PASS
% CD4+	54.9	55.0	0.0	PASS
CD3+ Abs Cnt	1497	1643	9.7	PASS
CD4+ Abs Cnt	823	903	9.8	PASS
CD8+ Abs Cnt	629	664	5.6	PASS
CD45+ Abs Cnt	2081	2309	11.0	PASS

#### Pass Criterion:

The pass criterion is a less than 15% difference between the blood lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

#### Comments:

3 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager



Date: 05 May 2022