

# TransFix<sup>®</sup>

## Certificate of Analysis

### TransFix<sup>®</sup>/EDTA Vacuum Blood Collection Tubes (TVT)

**RUO** **REF** : TVT-03-1-RUO, TVT-03-2-RUO, TVT-03-50-RUO **LOT** : 30071

 : 2026-09-30

Start Date of Manufacture: **2025-03-13**

TransFix Batch Number used in TVT manufacture: **30067**

Start Date of Quality Control Testing: **2025-04-17**

	A	B		
	Control Day 0	TVT Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	80.4	77.6	3.5	PASS
% CD4+	57.0	57.6	1.1	PASS
CD3+ Abs Cnt	1574	1604	1.9	PASS
CD4+ Abs Cnt	897	924	3.0	PASS
CD8+ Abs Cnt	620	641	3.4	PASS
CD45+ Abs Cnt	2002	2090	4.4	PASS

#### Pass Criterion:

The pass criterion is a less than 15% difference between the blood lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

#### Comments:

3 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager

*N Woodham*

Date: 30 Apr 2025