

# TransFix<sup>®</sup>

## Certificate of Analysis TransFix<sup>®</sup> Cellular Antigen Stabilisation Reagent

**LOT** : 30028

: 2025-08-31

Start Date of Manufacture: **2023-08-21**Start Date of Quality Control Testing: **2023-09-14**

	A	B		
	Control Day 0	TransFix Day 11	% Diff (A-B)/A	Pass / Fail
% CD3+	78.5	80.5	2.6	PASS
% CD4+	56.9	57.8	1.6	PASS
CD3+ Abs Cnt	3578	3324	7.1	PASS
CD4+ Abs Cnt	2034	1922	5.5	PASS
CD8+ Abs Cnt	1393	1285	7.8	PASS
CD45+ Abs Cnt	4667	4211	9.8	PASS

**Pass Criterion:**

The pass criterion is a less than 15% difference between the lymphocyte absolute count (CD45+ Abs Cnt), T cell absolute count and % (CD3+), T helper cell absolute count and % (CD4+) and cytotoxic T cell absolute count (CD8+) on day 0 and day 11.

**Comments:**

3 separately stained tests were performed at each of the time points. These results were averaged and all parameters passed the acceptance criterion.

Signature of Quality Control Manager



Date: 27 Sep 2023